



Certificate of Traceability

Manufacturer Certification

This certifies that **Boditech Med Inc.**, using **ichroma HbA1c Test** has participated in and successfully completed the NGSP certification for manufacturers and is traceable to the **Diabetes Control and Complications Trial Reference method**. The comparison was performed with: **Reference Material Institute for Clinical Chemistry Standards (ReCCS) ASRL#1**

The system evaluated was:

Instrument: ichroma Reader	Calibrator Lot: 130013A, 120355A, 415501A
Reagent Lot: AAKYA03.1	Calibrator Assigned Values: 5.1%, 5.80%, 6.5%, 8.1%, 11.2%, 15.2%

Date of Certification: January 1, 2015

Certification Expires: January 1, 2016

Daniel South

NGSP Steering Committee Chair

Randee R Little PhD.

NGSP Network Coordinator

Massao Umemoto, Ph.D.

SRL director/ supervisor